

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

<p>Product Produit</p> <p>Name and address of the applicant Nom et adresse du demandeur</p> <p>Name and address of the manufacturer Nom et adresse du fabricant</p> <p>Name and address of the factory Nom et adresse de l'usine</p> <p><small>Note: When more than one factory, please report on page 2 Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{eme} page</small></p> <p>Ratings and principal characteristics Valeurs nominales et caractéristiques principales</p> <p>Trademark (if any) Marque de fabrique (si elle existe)</p> <p>Type of Manufacturer's Testing Laboratories used Type de programme du laboratoire d'essais constructeur</p> <p>Model / Type Ref. Ref. De type</p> <p>Additional information (if necessary may also be reported on page 2) Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2^{eme} page</p> <p>A sample of the product was tested and found to be in conformity with Un échantillon de ce produit a été essayé et a été considéré conforme à la</p> <p>As shown in the Test Report Ref. No. which forms part of this Certificate Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat</p>	<p>Power Supply</p> <p>SL POWER ELECTRONICS CORP BLDG A 6050 KING ST VENTURA CA 93003, USA</p> <p>SL POWER ELECTRONICS CORP BLDG A 6050 KING ST VENTURA CA 93003, USA</p> <p>SL XIANGHE POWER ELECTRONICS CORP NO 4 SHUANGXING NORTH RD XINGHEÂ ECONOMIC & TECHNOLOGICAL DEVELOPMENT ZONE XIANGHE HEBEI CHINA</p> <p><input type="checkbox"/> Additional Information on page 2 See Page 2</p>  <p>SMT</p> <p>MINT1110AXYKZ Series See Page 2</p> <p><input checked="" type="checkbox"/> Additional Information on page 2</p> <p>IEC 60601-1(ed.3)</p> <p>E116994-A78-CB-1 issued on 2012-03-30</p>
---	---

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2012-04-22

Signature:



Jolanta M. Wroblewska



Model Details:

MINT1110AXYKZ Series, where MINT is Medical Internal Model, 1 is the number of outputs 110 is the output wattage i.e. 110 W, A: is any Alpha character A-Z which indicates changes not related to Safety X: is a 2 digit numeric character that represents the Output Voltage. Voltage Range is 12-24Vdc. Y: Output Connector Options K: Input Connector Options Z: Configuration: 01 is Standard, 02-99 Modifications which have no impact on Safety Model
MINT1110A1808K02

Ratings:

Input: 100-240V~, 50-60Hz, 1.4A
Output: MINT1110A1208K01 Series - Rated 12Vdc, 7.5A
MINT1110A1508K01 Series - Rated 15Vdc, 6.5A
MINT1110A1808K01 Series - Rated 18Vdc, 5.8A
MINT1110A1808K02 Series - Rated 18Vdc, 5.8A
MINT1110A1908K01 Series - Rated 19Vdc, 5.8A
MINT1110A2408K01 Series - Rated 24Vdc, 4.6A

Additional Information:

Additionally evaluated to EN 60601-1:2006; National Differences specified in the CB Test Report.

Additional information (if necessary)**Information complémentaire (si nécessaire)**

UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2012-04-02

Signature:

Jolanta M. Wroblewska