

DEMKO CERTIFICATE

Certificate No. D-03313-A3
Page 1/3
Date of Issue 2015-03-03

Certificate Holder SL POWER ELECTRONICS CORP
BLDG A
6050 KING DR
VENTURA, CA 93003 USA

Manufacturer SL POWER ELECTRONICS CORP
BLDG A
6050 KING DR
VENTURA, CA 93003 USA

Production site INDUSTRIAS S L S A DE C V
CIRCUITO SIGLO XXI 2055
COL PARQUE INDUSTRIAL EX-XXI
MEXICALI, BC 21254 Mexico

See Page 2

Certified Product Model Switching Power Supply
MB65SXXWYY
See Page 2

Trademark



Rated Voltage / Frequency 100-240V~, 50-60Hz

Rated Current / Power 1.5A

Insulation Class MB65SXXKYY series (Class I type), and MB65SXXCY series (Class II type)

Degree of protection (IP) IPX0

Tested acc. to EN 60601-1:2006/A11:2011, EN 60601-1:2006/A12:2014, EN 60601-1:2006

Test Report No. E116994-A104-CB-1 issued on 2015-03-03

Additional For building-in

Expire date 2018-12-24

Certification Manager
Jan-Erik Storgaard

This is to certify that representative sample(s) of the Product described herein ("Certified Product") have been investigated and found in compliance with the Standard(s) indicated on this Certificate, in accordance with the D Mark Requirements. The Designated Certificate holder is entitled to use the d or for cables <DEMKO> for the Certified Product manufactured at the production site(s) identified above, in accordance with the D Mark Service Agreement, including without limitation the D Mark Testing and Certification Services Service Terms. Only those Products bearing the D Mark should be considered as being covered by UL's D Mark Service. This Certificate shall remain valid through the expiration date, unless terminated earlier in accordance with the Service Agreement including without limitation if the Standard identified on this Certificate is amended or withdrawn prior to the expiration date.

Certification Body UL International Demko A/S, Borupvang 5A, DK-2750
Ballerup, Denmark, Tel. +45 44 85 65 65,
info.dk@ul.com
www.ul-europe.com



Appendix DEMKO CERTIFICATE

Certificate No. D-03313-A3
Page 2/3
Date of Issue 2015-03-03

Model Details:

MB65SXXWYY Where XX represents output voltage which may be any number from 12 to 48, W can be K for Class I construction and C for Class II construction or blank, YY may be any number from 00 to 99 or blank, designates additional configurations indicating non-safety related options.

Production Sites:

SL XIANGHE POWER ELECTRONICS CORP
No.B-02-03,North side of Landscape Ave, Qibu District,
Environmental Industrial Park
Xianghe, 065400 Hebei China

Additional Information:

This certificate replaces the certificate no. D-03313-A2 issued 2014-10-08 due to:

1. Delete factory located in Hebei China
2. Add EN 60601-1/A12:2014 to the evaluation.

Output rating @ ambient temperature condition: $T_{ambient} \leq 50$ degree C

MB65S12K:12Vdc/5.4A
MB65S15K:15Vdc/4.33A
MB65S18K:18Vdc/3.6A
MB65S24K:24Vdc/2.7A
MB65S28K:28Vdc/2.32A
MB65S32K:32Vdc/2.03A
MB65S48K:48Vdc/1.35A

Output rating @ ambient temperature condition: $T_{ambient} \leq 70$ degree C

MB65S12K:12Vdc/3.8A
MB65S15K:15Vdc/3.04A
MB65S18K:18Vdc/2.53A
MB65S24K:24Vdc/1.9A
MB65S28K:28Vdc/1.63A
MB65S32K:32Vdc/1.42A
MB65S48K:48Vdc/0.95A


Certification Body

UL International Demko A/S, Borupvang 5A, DK-2750
Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@ul.com
www.ul-europe.com



Appendix DEMKO CERTIFICATE

Certificate No. D-03313-A3
Page 3/3
Date of Issue 2015-03-03



Certification Body

UL International Demko A/S, Borupvang 5A, DK-2750
Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@ul.com
www.ul-europe.com

