


IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

CB TEST CERTIFICATE

Product	Switched Mode Power Supply
Name and address of the applicant	Astec International Ltd 16TH FLLU PLAZA2 WING YIP STKWUN TONG KOWLOON HongKong
Name and address of the manufacturer	EXCELSYS TECHNOLOGIES LTD UNIT 27EASTGATE BUSINESS PARK LITTLE ISLAND CORK IRELAND
Name and address of the factory	SHENZHEN WATT ELECTRONICS CO. LTD NO. 5 TUNNEL 1, TANGFANG GARDEN, 35 DISTRICT, BAOAN 518101 SHENZHEN, GUANGDONG CHINA
Note: When more than one factory, please report on page 2	<input type="checkbox"/> Additional Information on page 2
Ratings and principal characteristics	Input: 100-240V, 50-60Hz, 8A-4.7A, Class I, no applied part. Output: 1000W maximum (see Table 2, Table 3, Table 4 in General Information in Test Report).
Trademark / Brand (if any)	
Type of Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	CX10M-000000-defgh (CoolX CoolPac without CoolMod), CX10M-uvwxyz-defgh (CoolX CoolPac with CoolMod)
Additional information (if necessary may also be reported on page 2)	The report was revised to include technical modifications. <input checked="" type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005
As shown in the Test Report Ref. No. which forms part of this Certificate	E182560-D1007-1/A1/C0-CB issued on 2020-12-28

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnamesDate: 2021-01-07
Original Issue Date: 2020-11-17

Signature:

Jan-Erik Storgaard

Additional Information:

The original report was modified to include following changes/additions:
- update of LoCC.

National Differences specified in the CB Test Report.

See Test Report for Part Numbering System Configurations.

The Risk Management requirements of the standard were not addressed.

Additional information (if necessary)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/incbnames

Date: 2021-01-07

Original Issue Date: 2020-11-17

Signature:

Jan-Erik Storgaard