



Ref. Certif. No.

**US-18047-UL**

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

**CB TEST CERTIFICATE**

**CERTIFICAT D'ESSAI OC**

Product  
Produit

Component – Switching Power Supply

Name and address of the applicant  
Nom et adresse du demandeur

SL POWER ELECTRONICS CORP  
6050 KING DRVIE BLDG. A  
VENTURA CA 93003, USA

Name and address of the manufacturer  
Nom et adresse du fabricant

SL POWER ELECTRONICS CORP  
6050 KING DRVIE BLDG. A  
VENTURA CA 93003, USA

Name and address of the factory  
Nom et adresse de l'usine

See Page 2

Note: When more than one factory, please report on page 2  
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2<sup>ème</sup> page

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

Input: 100-240V~, 50-60Hz, 3.0A  
Output: See Enclosure, Miscellaneous, Marking Plate for output ratings.

Trademark (if any)  
Marque de fabrique (si elle existe)



Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais constructeur  
Model / Type Ref.  
Ref. De type

SMT

MINT1200AXXYKZZ, MINT1180AXXYKZZ, See Page 2

Additional information (if necessary may also be reported on page 2)  
Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2<sup>ème</sup> page

This CB Test Report comprises 2 Enclosures.  
See Test Report for National Differences.

A sample of the product was tested and found to be in conformity with  
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1(ed.3)

As shown in the Test Report Ref. No. which forms part of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

11CA42173 issued on 2011-11-09

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



Underwriters Laboratories Inc. / GMA Certification Department, US  
333 Pflingsten Road, Northbrook, IL 60062-2096  
United States of America  
TEL INT\* +1 847 664 3008, FAX INT\* +1 847 313 3008  
email: jolanta.m.wroblewska@us.ul.com

Signature:

Jolanta M. Wroblewska

Date: 2011-11-10



Ref. Certif. No.

**US-18047-UL**

Model Details:

MINT1200AXXYKZZ, MINT1180AXXYKZZ;  
where A is A (Class I construction) or B (Class II construction), where XX is any number 12 through 48, where YY or ZZ is any number 01 through 99, designates additional configurations indicating non-safety related options.

Factories:

SL XIANGHE POWER ELECTRONICS CORP  
NO.4, SHUANGXING NORTH ROAD  
XIANGHE ECONOMIC AND TECHNOLOGICAL DEVELOPMENT ZONE  
XIANGHE COUNTY HEBEI PROVINCE, CHINA

**Additional information (if necessary)**  
**Information complémentaire (si nécessaire)**



Date: 2011-11-10

Underwriters Laboratories Inc. / GMA Certification Department, US  
333 Pfingsten Road, Northbrook, IL 60062-2096  
United States of America  
TEL INT\* +1 847 664 3008, FAX INT\* +1 847 313 3008  
email: jolanta.m.wroblewska@us.ul.com

Signature:

Jolanta M. Wroblewska