

COVER PAGE FOR TEST REPORT

Product Category:	Power Supplies, Medical and Dental
Product Category CCN:	QQHM2, QQHM8
Test Procedure:	Component Recognition
Product:	AC-DC Adaptor
Model/Type Reference:	<p>BP(1)050(2)(3)(4)(5);</p> <p>(1) - Family Related Designs: A to Z</p> <p>(2) - Output: S (S=Single)</p> <p>(3) - Output Voltage: 05, 06, 09, 12, 15, 16, 18, 24, 48</p> <p>(4) - Standard Input Cord Options:</p> <p>F : (Class I = IEC320-C14)</p> <p>Q: (Class II = IEC320-C18)</p> <p>N: ((Class II = IEC320-C8)</p> <p>(5) - Custom Options (Marking, Cord etc.): Number 00 to 99</p> <p>(1)ENB1050(2)(3)(4)(5)(6);</p> <p>(1) - Family Related Designs: A to Z</p> <p>(2) - Design Revision Changes: A to Z (Standard)</p> <p>(3) - Output Voltage: 05, 06, 09, 12, 15, 16, 18, 24, 48</p> <p>(4) - Standards Output Cord Options: Number 00 to 99</p> <p>(5) - Standard Input Cord Options:</p> <p>F: (Class I = IEC320-C14)</p> <p>Q: (Class II = IEC320-C18)</p> <p>N: ((Class II = IEC320-C8)</p> <p>(6) - Custom Options (Marking, Cord etc.): Number 00 to 99</p> <p>BP(1)060(2)(3)(4)(5);</p> <p>(1) - Family Related Designs: A to Z</p> <p>(2) - Output: S (S=Single)</p> <p>(3) - Output Voltage: 05, 06, 09, 12, 15, 16, 18, 24, 48</p> <p>(4) - Standard Input Cord Options:</p> <p>F : (Class I = IEC320-C14),</p> <p>Q: (Class II = IEC320-C18),</p> <p>N: ((Class II = IEC320-C8)</p> <p>(5) - Custom Options (Marking, Cord etc.): Number 00 to 99</p> <p>(1)ENB1060(2)(3)(4)(5)(6);</p> <p>(1) - Family Related Designs: A to Z</p> <p>(2) - Design Revision Changes: A to Z (Standard)</p> <p>(3) - Output Voltage: 05, 06, 09, 12, 15, 16, 18, 24, 48</p> <p>(4) - Standards Output Cord Options: Number 00 to 99</p> <p>(5) - Standard Input Cord Options,</p> <p>F : (Class I = IEC320-C14),</p> <p>Q: (Class II = IEC320-C18),</p> <p>N: ((Class II = IEC320-C8)</p> <p>(6) - Custom Options (Marking, Cord etc.), Number : 00 to 99</p>
Rating(s):	<p>For Model BP(1)050(2)(3)(4)(5) and (1)ENB1050(2)(3)(4)(5)(6)</p> <p>- Rated Input: 100-240 Vac, 50-60Hz, 1.5 A.</p> <p>- Rated Output: +5Vdc/6A or +6Vdc/5A or 9Vdc/5A or 12Vdc/4.2A,</p>

	<p>15Vdc/3.36A, 16Vdc/3.15A or 18Vdc/2.8A or 24Vdc/2.1A</p> <p>For Model BP(1)060(2)(3)(4)(5) and (1)ENB1060(2)(3)(4)(5)(6)</p> <ul style="list-style-type: none">- Rated Input: 100-240 Vac, 50-60Hz, 1.5 A.- Rated Output: +5Vdc/7A or +6Vdc/6A or 9Vdc/6A or 12Vdc/5A, 15Vdc/4A, 16Vdc/3.7A or 18Vdc/3.4A or 24Vdc/2.7A <p>Standards:</p> <p>UL 60601-1, 1st Edition, 2006-04-26 (Medical Electrical Equipment, Part 1: General Requirements for Safety)</p> <p>CAN/CSA-C22.2 No. 601.1-M90, 2005 (Medical Electrical Equipment - Part 1: General Requirements for Safety)</p> <p>Applicant Name and Address:</p> <p>BRIDGEPOWER CORP 964 GOSAEK-DONG GWONSEON-GU SUWON-SI GYEONGGI-DO 441-813 KOREA</p>
	<p>This Report includes the following parts, in addition to this cover page:</p> <ol style="list-style-type: none">1. Specific Inspection Criteria2. Specific Technical Criteria3. Clause Verdicts4. Critical Components5. Test Results6. National Differences7. Enclosures

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of Underwriters Laboratories Inc. ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.


UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability.

Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL.

Test Report By:


DongHa Lee
Assistant Project Handler
UL Korea, Ltd.

Reviewed By:


ChanYo Won
Section Manager
UL Korea, Ltd.