



Product Service

CERTIFICATE

No. Z2 17 03 13890 02807

Holder of Certificate: **Astec International Ltd.**
 16th Floor, Lu Plaza, 2 Wing Yip Street
 Kwun Tong
 Kowloon
 HONG KONG



Certification Mark:



Product: **Switching power supply unit
 (Switching Power Supply)**

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 682701700101

Valid until: 2019-06-19

Date, 2017-03-20

(Yager Bi)



Page 1 of 2

**CERTIFICATE****No. Z2 17 03 13890 02807****Model(s):** DA10-050EU-M, DA10-050UK-M,
DA10-050MP-M**Parameters:**
Rated Input : 100-240V, 50/60Hz, 0.3A Max.
Rated Output : 5VDC, 2.0A Max., 10W Max.
Protection Class : II
Degree of Protection : IP40 for model DA10-050MP-M;
IP42 for other models**Remarks:**

- The maximum operating temperature is 40°C.
- Clearance was evaluated for operating altitude up to 3000m above sea level.
- These power supplies have been evaluated according to EN 60601-1:2006/A1:2013 with the following conditions:
 - 1) The output was not evaluated as patient connected circuits.
 - 2) Compliance with the requirements for EMC shall be evaluated for the end use product.
 - 3) These power supplies have been investigated only as a component part for use in equipment where the suitability of the combination is subject to end product investigation.
 - 4) These power supplies must be installed in accordance with the instruction manual.
 - 5) The leakage current test shall be checked in end product.
 - 6) The risk management requirements of the standard were not addressed.
 - 7) Clearance/creepage distance and dielectric strength were evaluated and fulfilled the requirements for MOPP.

Tested according to: EN 60601-1:2006/A1:2013
EN 60950-1:2006/A2:2013**Production Facility(ies):** 28532, 62777