



Ref. Certif. No.

US-38113-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product Component switch mode power supply for use in medical devices

Name and address of the applicant EXCELSYS TECHNOLOGIES LTD
UNIT 27
EASTGATE BUSINESS PARK LITTLE ISLAND CO CORK
IRELAND

Name and address of the manufacturer EXCELSYS TECHNOLOGIES LTD
UNIT 27
EASTGATE BUSINESS PARK LITTLE ISLAND CO CORK
IRELAND

Name and address of the factory SHENZHEN WATT ELECTRONICS CO LTD
BAOAN BRANCH, 5 TUNNEL 1
TANGFANG GARDEN 35 DISTRICT
BAO'AN SHENZHEN, GUANGDONG, 518101
CHINA

Note: When more than one factory, please report on page 2

Additional Information on page 2

Ratings and principal characteristics AC Input: 200V-240V~, Amps: 16A-15A, Freq.: 50-60Hz
Maximum Total Output Power: 3000 W max

Trademark (if any)

Customer's Testing Facility (CTF) Stage used CTF Stage 2

Model / Type Ref. C3a00000000000bcd (3000W CoolX CoolPac front-end),
C3ajklmnpvwxzbcdbcd (3000W CoolX Configured Power Supply)
Refer to Test Report - Miscellaneous enclosure for nomenclature.

Additional information (if necessary may also be reported on page 2) Additional Information on page 2

A sample of the product was tested and found to be in conformity with IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005

As shown in the Test Report Ref. No. which forms part of this Certificate E230761-D1003-1/A0/C0-CB issued on 2020-12-23

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2021-05-14

Signature: Jolanta M. Wroblewska



Ref. Certif. No.

US-38113-UL

Additionally evaluated to:

EN 60601-1:2006 / A1:2013 / A12:2014; National Differences specified in the CB Test Report.

The product has been evaluated with the exception of Risk Management assessment.

Additional information (if necessary)



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Signature:

Jolanta M. Wroblewska

Date: 2021-05-14